

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02640	Calib. Date: 05/27/2025
Certification File No.:	02641	Calib. No.: 00050
Linearity File No.:	02642	Cert. Date: 05/27/2025
Solution File No.:	02657	Cert. No.: 00039
Sequential File No.:	02657	Lin. Date: 05/27/2025
		Lin. No.: 00039
		Soln. Date: 07/23/2025
		Soln. No.: 00472
		File Date: 07/23/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	24260	Expires: 07/23/2026
		Bottle No.: 0815

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:26D		
Control 1 EC	0.097%	09:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28D		
Control 2 EC	0.097%	09:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		
Control 3 EC	0.098%	09:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

[Handwritten Signature] *364

Badge No.: 364

Date: 07/23/2025